

## **EU DECLARATION OF CONFORMITY**

This declaration of conformity is issued under the sole responsibility of the

Manufacturer:

ZPA Nová Paka, a.s., , registered in the Commercial Register at the Regional

Court in Hradec Králové, Section B, Inlet 644. Registration No. 46504826

Seat:

Pražská 470, 509 39 Nová Paka, Czech Republic

Name:

The sensor relative and absolute pressure and vacuum HART INPRES 04

Type:

704

Basic data about product:

Type of supply mains: DC 24 V

The sensor is designed for the process industry for accurate measurement of absolute and relative pressure and vacuum of non-aggressive and aggressive liquids, gases and vapors in areas such as .: mining of raw materials, chemical and petrochemical industry, power industry, heavy industry.

The sensor is designed according to ČSN EN 61140 ed. 2 as electrical equipment of protection class III for the application in networks with the category specified impulse withstand voltage II and pollution grade 2 according to ČSN EN 61010-1 ed. 2 up (evaluation) device shall comply with Art. 6.3 of this standard.

## The above described object of the declaration is in conformity with the relevant EU harmonization legislation:

 Directive 2014/30/EU of the European Parliament and of the Council and standards: ČSN EN 61326-1 ed.2:2013, CISPR11, art. 6.2.1.3 and art. 6.2.2.3
ČSN EN 61000-4-2/B, ČSN EN 61000-4-3/A, ČSN EN 61000-4-4/B, ČSN EN 61000-4-5/B, ČSN EN 61000-4-6/A, EN 61000-4-8/A

## and declares that

properties of the product comply with requirements of basic safety principles and requirements of technical regulations and the product is, subjected to the conditions of correct installation, maintenance, operation and application, safe for the specified purpose; measures have been taken to ensure compliance of all marketed products with technical documentation and basic requirements of said Directive.

Last two digits of the year, in which the CE mark was placed on the product: 16 Signed for and on behalf of:

Nová Paka, April 20th, 2016

Name:

Petr Fól

Position:

Quality Manager

Signature:



